



Certificate of Conformity

VWR ZIP BAGS, SINGLE POCKET WITH WRITING FIELD

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **129-0325**

Lot No.: **032113**

L x W : **200mm x 300mm**

Pack : **1.000**

Material: **LDPE C150Y+LLDPE 218W (Virgin)**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications.**

Sterility : **Product labeled as sterile is gamma irradiated according to ISO 11137:2006, recommended practices in effect at the time of validation. Products labeled sterile meets a minimum requirement of 10⁻⁶ SAL, with the dose range of: Min.20KGY Max. 25 KGY**

Date of Sterilization : **15th Apr 2013**

Expiry Date : **15th Apr 2016**

This document has been produced electronically and is valid without a signature

Date of issue: **September 2013**