

Certificate of Conformity

VWR HDPE BOTTLE 60 ml, AMBER, WIDE MOUTH WITH SCREW CAP

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **215-3444**

Lot No: **V181011**

Material: **100% Virgin HDPE**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications**

Leakage test: **Ok [No leakage found]**

Compliant to: **FDA**

This document has been produced electronically and is valid without a signature

Date of issue: **January 2014**