Certificate of Conformity

VWR PLAIN BLENDER BAGS STERILE

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les specifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Cat. No.	Description	Capacity (ml)	Height (mm)	Width (mm)	Thickness (mm)	Pk
129-1102	Plain Blender bags sterile	400	300±3	175±3	0,07	20 bags of 25 pcs
129-1103	Plain Blender bags sterile	400	300±3	175±3	0,07	10 bags of 50 pcs

Material: Polyolefin complex (PE)

Food contact : Approve for food contact

Manufactured under ISO 9001:2008

Sterilization

Is made in compliance with to the GMP of quality Assurance and attest treatment by gamma rays between 5 to 12 Kilo Gray

Blender test : Bags have to be resistant for 3 minutes blending (frequency: 4 times / second)

Use temperature main body : 40°C / +80°C

Expiry: 10 years

Physical Requirements

Flexible, good mechanical resistance.
Can be frozen
Red sticker on every box serves as qualitive indicator of gamma ray processing.

Disclaimer: *VWR wants to state that these specifications can only be regarded as general information and will not discharge the user from his obligation to ensure the product is suitable for the intended use - If there is any doubt ask us.*

Date of issue: **November 2015**