Certificate of Conformity

VWR HDPE BOTTLE 30 ml, AMBER, WIDE MOUTH WITH SCREW CAP

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No : **215-3443**

Lot No: **V100709**

Specifications	Req. Values	Tolerance	Quality Control
External body diam.	34.5 mm	±0.50mm	**OK**
External Neck Diam	24.6 mm	±0.20mm	**OK**
Internal Neck Dimension	21.6 mm	±0.20mm	**OK**
Height	60 mm	±0.50mm	**OK**
Weight w/o cap	9.5 gm	±0.50gm	**OK**

Material : **Bottle: 100% Virgin HDPE**
Screw Cap: Polypropylene

QC Testing**: Representative production samples are collected and inspected in accordance with current applicable product specifications**

Leakage test at 635 mm Hg: **Ok**

Compliant to: **FDA 21CFR 177.1520**
IS 10146 – 1982:Specification of Polyethylene for it safe use in contact with food stuff and drinking water
IS 10141 – 1982: Positive list of constitutes of Polyethylene for its safe use in contact with food stuff and drinking water.

This document has been produced electronically and is valid without a signature

Date of issue: **March 2010**