



Product Specifications sheet

VWR LABORATORY BOTTLES, NARROW MOUTH, WITH SCREW CAP

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Cat. N°	Diameter (mm)	Height (mm)	Capacity (ml)
215-0057	186	330	5000
215-0058	234	410	10000
215-0059	299	505	20000
215-1592	56	100	100
215-1593	70	138	250
215-1594	86	176	500
215-1595	101	225	1000
215-1596	136	260	2000

Intended uses: Laboratory applications

Material: Borosilicate 3.3 glass

Shape: Round, with graduations, DIN GL 45 thread, pouring ring and blue PP screw cap

Compliant to: ISO 3585 Borosilicate glass 3.3 - properties
ASTM E438-92 Type 1
Eur. Pharmacopoeia 5.0, glass type I



Product Specifications sheet

VWR LABORATORY BOTTLES, NARROW MOUTH, WITH SCREW CAP

Chemical Properties: Have been tested by the accredited Certification body Nr 3019 Glass Institute Hradec Kralovéd and found compliant to CSN EN ISO 3585, art.4.1,4.2,4.3a, 4 and ASTM E438-92.)
Hydrolytic class HGB 1 according to ISO 719
Hydrolytic class HGA 1 according to ISO 720

Disclaimer: VWR wants to state that these specifications can only be regarded as general information and will not discharge the user from his obligation to ensure the product is suitable for the intended use - If there's any doubt ask us.

Date of issue: **February 2013**

VWR International bvba/sprl, Haasrode Research Park Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium

<http://www.vwr.com>



Version: 2.0