Certificate of Conformity

VWR BLENDER BAG WITH LATERAL FILTER

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **129-0730**

Lot No: **130419**

Capacity (ml)	W×L (mm)	Packed	Pk
400	**300 x 190**	**60 bags of 25**	**1.500**

Material: **100% Virgin raw material (PA+PE+Non-woven polyester), no recycled material allowed**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications.**

Blending Test: **Bags have to be resistant for 3 minutes blending (frequency: 4times/sec)**

Sterility: **Products labeled sterile meets a minimum requirement of 10^{-3} SAL.**

Date of Sterilization: **10th Jun 2013**

Expiry Date: **10th Jun 2016**

This document has been produced electronically and is valid without a signature

Date of Issue: **February 2014**