



Certificate of Conformity

BOTTLE PP 125 ml ROUND NARROW MOUTH

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No : **215-3428**

Lot No: **V010109**

Specifications	Req. Values	Tolerance	Quality Control
External body diam.	51.5 mm	±0.50mm	OK
External Neck Diam	21 mm	±0.20mm	OK
Internal Neck Dimension	17.5 mm	±0.20mm	OK
Height	98.5 mm	±0.50mm	OK
Weight w/o cap	19 gm	±0.50gm	OK

Material : **100% Virgin PP**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications**

Leakage test at 635 mm Hg: **Ok**

Autoclaving at 121°C, 15 psi, 20min: **Ok**

Compliant to: **FDA 21CFR 177.1520**

IS 10146 – 1982: Specification of Polyethylene for its safe use in contact with food stuff and drinking water

IS 10141 – 1982: Positive list of constituents of Polyethylene for its safe use in contact with food stuff and drinking water.

This document has been produced electronically and is valid without a signature

Date of issue: **March 2010**