



Certificate of Conformity

VWR HIGH PERFORMANCE BOTTLES, 30 ML, WIDE MOUTH, WITH SCREW CAP

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No : **215-3449**

Lot No: **V161109**

Specifications	Req. Values	Tolerance	Quality Control
Thread – Dim	27.3	±0.20 mm	Ok
External Neck - Dim	24.6	±0.20 mm	Ok
Neck Inner - Dim	21.1	±0.20 mm	Ok
Outer Body - Dia	34.5	±0.50 mm	Ok
Height	60	±0.50 mm	Ok
Wall Thickness LHS(UP/DN)	1.3	±0.20 mm	Ok
Wall Thickness RHS(UP/DN)	1.3	±0.20 mm	Ok
Weight W/O Cap	9.5	±0.50 mm	Ok
Closer Dia	28	±0.20 mm	Ok

Material: **100% Virgin Polypropylene and FDA approved**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications**

Leakage test at 635 mm Hg: **Ok**

Autoclaving : **Ok**

This document has been produced electronically and is valid without a signature

Date of Issue: August, 2011