

Declaration of conformity

We: **VWR International Belgium BVBA**

Geldenaaksebaan 464

B-3001 Leuven

Belgium

Tel.: +32 (0) 16 385 011

Fax: +32 (0) 16 385385

E-mail: info@be.vwr.com

hereby declare under our sole responsibility that the products:

**Horizontal Gel Electrophoresis Systems: 700-7005 (ATG1), 700-7031(ATG3), 700-7051 (ATG2),
700-7166(ATG4), 700-7213 (MAXI VARIGEL).**

Iso-Electric Focussing Horizontal Gel Electrophoresis System: 700-7121.

**Vertical Gel Electrophoresis Systems: 700-7271, 700-7276, 700-7273, 700-7278, 700-7272, 700-7277,
700-7274, 700-7279.**

Denaturing Gradient Gel Electrophoresis System + Caster: 700-7149.

Mini Electro-blotting Units: 700-7130 (BTV400), 700-7131 (BTV 100).

Maxi Electro-blotting Modules: 700-7282 (for CTV100), 700-7283 (for CTV400).

Maxi Semi-Dry Blotting Unit: 700-7156 (V20-SDB).

Vacuum Heated Gel Drying Units: 700-7124(MGD-4534), 700-7125(MDG-5040).

to which this declaration relates, are in conformity with the following Council directives of the EU Member states:

Council directive on harmonization of the laws of the Member states:
--

73/23/EEC: relating to electrical equipment for use within certain voltage limits + Amendments. 89/336/EEC: relating to Electromagnetic Compatibility + Amendments.
--

And meet the essential requirements of the mentioned directives.

This equipment has been tested according to following standards:

Titles of the applied standards	References numbers
Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: general requirements.	EN 61010-1:1993 IEC 1010-1:1990 + Amendment (1992)

The technical documentation required to demonstrate that the products meet the requirements of theses directives has been compiled and can be made available for inspection by the relevant enforcement authorities.

Date of original issue: 15.06.2007

Signature:



Raoul Biskupek
Managing Director